OXFORD LAB FINE CHEM LLP
ISO 9001-2008 Certified Company

Regd Office: Unit no 12, 1st Floor,
Neminath Industrial Estate No.6,
Navghar, Vasai (East), Palghar - 410210.
Maharashtra, INDIA.

Tel: +91 250 2390032 / 2390989 / 2390990
Email: sales@oxfordlabchem.com /
info@oxfordlabchem.com
Web: www.oxfordlabchem.com

Oxford
Range of
Laboratory Chemicals

TECHNICAL DATA SHEET

MacConkey Broth W/BCP and NaCl

Principle

MacConkey broth purple is composed of peptone, lactose, sodium taurocholate, sodium chloride and bromocresol purple. Peptone provides nitrogen and other nutrients necessary for the growth of microorganism. Lactose is a carbon source and plays an important role for selection of lactose fermenting microbes. Sodium taurocholate is selective agents, inhibit growth of gram-positive organisms. Bromocresol purple is pH indicator dye. Sodium chloride maintains osmatic equilibrium. A lactosefermenting organism shows prominent growth and produce acid and gas, causing the medium to turn yellow. Non-fermenting organisms produce good growth but will not produce acid or gas.

Use: For the presumptive identification of coliforms from water, milk and foods products.

Contents*

Ingredients	Gram/Litre
Peptone	20.000
Lactose	10.000
Sodium taurocholate	5.000
Sodium chloride	5.000
Bromocresol Purple	0.010
pH at 25°C	7.4 ±0.2

*** Formula adjusted for optimum performance and parameters**

Directions: Dissolve 40.00 grams in 1000 ml distilled water, boil to dissolve the medium completely and distribute in test tubes with inverted Durham tube. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 min, cool it to 42-45 °C and inoculate test sample aseptically.

Specimens' types analyzed

Food, dairy and water samples, pharmaceutical samples, clinical and non-clinical samples. etc.

OXFORD LAB FINE CHEM LLP
ISO 9001-2008 Certified Company

Regd Office: Unit no 12, 1st Floor,
Neminath Industrial Estate No.6,
Navghar, Vasai (East), Palghar - 410210.
Maharashtra, INDIA.

Tel: +91 250 2390032 / 2390989 / 2390990
Email: sales@oxfordlabchem.com /
info@oxfordlabchem.com
Web: www.oxfordlabchem.com

Oxford
Range of
Laboratory Chemicals

Precautions to be taken

These microbial media are intended for the in-vitro use only. All the handling, experiments, storage, and discarding should be performed with the help of skilled and knowledgeable technicians and as per the established guidelines. The material should be disposed only after proper sterilization by autoclaving. Please go through the MSDS of the media to avoid any accidents or in emergency.

Performance and Evaluation

The expected performance of the medium is liable to use as per the direction on the label when stored at optimum conditions and within expiry date.

Quality Control

Appearance	**Light beige colored free flowing, homogeneous powder**
Reaction of 4.00% solution	**7.4 ±0.2 at 25 ˚C**
pH	**7.20- 7.60**
Color and clarity of ready medium	**Purple clear solution**
Growth Promotion properties	**Best at ≤ 100 CFU at 32-37 ºC for 18-72 h**
Indicative properties	**Optimum at ≤ 100 CFU at 32-37 ºC for 18-48 h**
Negative control	**Performed using sterile distilled water**

Different Microbial Response: Cultural characteristics observed after incubation at 42-44°C for 18-48 hours. Inoculum 50-100 CFU.

Organism	ATCC	Growth	Acid production	Gas production
Escherichia coli	8739	Luxuriant	Yellow color (Positive)	Positive
Staphylococcus aureus	25923	Inhibited	--	--

Cultural characteristics observed after incubation at 35 ± 2°C for 18-48 hours. Inoculum 50-100 CFU.

Organism	ATCC	Growth	Acid production	Gas production
Escherichia coli	8739	Luxuriant	Yellow color (Positive)	Positive
Klebsiella aerogenes	13048	Luxuriant	Yellow color (Positive)	Positive
Staphylococcus aureus	25923	Inhibited	--	--

OXFORD LAB FINE CHEM LLP
ISO 9001-2008 Certified Company

Regd Office: Unit no 12, 1st Floor,
Neminath Industrial Estate No.6,
Navghar, Vasai (East), Palghar - 410210.
Maharashtra, INDIA.

Tel: +91 250 2390032 / 2390989 / 2390990
Email: sales@oxfordlabchem.com /
info@oxfordlabchem.com
Web: www.oxfordlabchem.com

Oxford
Range of
Laboratory Chemicals

Storage and Shelf Life: **The product is highly hygroscopic; keep the container tightly closed at all times and store it properly as per the conditions mentioned on the label. The declared expiry is valid only when stored as per the conditions mentioned on the label. Note: Sterilize media immediately after reconstitution.**

Disposal: **To avoid the contamination or propagation of any hazardous microbes the used, unusable or modified preparation of this product must be disposed after autoclaving after completion of task.**

Reference

1. **Atlas, R. M. (2005). Handbook of media for environmental microbiology. CRC press.**
2. **Difco Manual (1998). 11th Edition. Difco Laboratories., Division of Becton Dickinson and Company, Sparks, Maryland, USA.**
3. **Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.**
4. **Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.**
5. **Rand, M. C., Arnold E. Greenberg, and Michael J. Taras, (1976), Standard methods for the examination of water and wastewater. Prepared and published jointly by American Public Health Association, American Water Works Association, and Water Pollution Control Federation.**

Disclaimer:

****

The information contained herein in good faith but makes no representations as to its comprehensiveness or accuracy. This document is intended only as a guide to the appropriate precautionary handling of the material by a properly trained person using this product. Individuals receiving the information must exercise their independent judgment in determining its appropriateness for a particular purpose.

Oxford Lab Fine Chem LLP makes no representations or warranties, either express or implied, including without limitation any warranties of merchantability, fitness for a particular purpose with respect to the information set forth herein or the product to which the information refers. Accordingly, Oxford Lab Fine Chem LLP will not be responsible for damages resulting from use of or reliance upon this information.